

Drinking Water Scheme – Preliminary Intended Results Notification

These results are provided for guidance and are determined from the examinations undertaken during quality control testing of the samples in the FEPTU laboratory. The enumeration results may show minor variations from the final results, which are calculated from the participants' consensus results, and included in the distribution report.

Distribution: **W218** Samples: **W218A and W218B**

Intended results:

Enumeration	W218A Median	W218B Median
Coliform bacteria per 100mL	60	99
<i>Escherichia coli</i> per 100mL	0	38
Enterococci per 100mL	0	58
<i>Pseudomonas aeruginosa</i> per 100mL	51	1.04x10 ²
<i>Clostridium perfringens</i> per 100mL	40	34
Colony count 37°C/48 h per mL	37	2
Colony count 22°C/72 h per mL	40	59

Sample contents:

W218A	<i>Enterobacter cloacae</i> , <i>P. aeruginosa</i> , <i>C. perfringens</i> , <i>Micrococcus</i> sp. and <i>Staphylococcus saprophyticus</i>
W218B	<i>Klebsiella pneumoniae</i> , <i>E. coli</i> , <i>Enterococcus hirae</i> , <i>P. aeruginosa</i> , <i>C. perfringens</i> and <i>Microbacterium</i> sp.

Laboratory identification numbers of participants who did not return results this distribution:
1093

The distribution report providing details of your individual performance and an analysis of all participants' results will be dispatched within three weeks. The report may also include sample specific comments.

Repeat samples are available for up to one year from date of this letter for participants who experienced problems with any of the examination. There is no need to wait until you have received the report to order repeat samples.

The next distribution of Drinking Water Scheme samples will be dispatched during the week beginning 17 February 2025.

If you require any technical advice or wish to make comments or suggestions relating to the scheme, please contact us.

Validated by:
Divya George
Healthcare Science
Specialist

Date: 10 February 2025



0006